



SEP 12 2003

Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

Mr. Peter Rumswinkel
VP/ General Manager
1025 Street James Church Road
PO Box 468
Newton, North Carolina 28658

Re: K032150
Trade/Device Name: Safety- Multifly
Regulation Number: 880.5440, 862.1675
Regulation Name: Intravascular Administration
Regulatory Class: II
Product Code: FPA, JKA
Dated: July 11, 2003
Received: July 14, 2003

Dear Mr. Rumswinkel:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Office of Compliance at (301) 594-1418. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its Internet address <http://www.fda.gov/cdrh/dsma/dsmamain.html>

Sincerely yours,



Susan Runner, DDS, MA
Interim Director
Division of Anesthesiology, General Hospital
Infection Control and Dental Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

SARSTEDT



Sarstedt, Inc. - P.O. Box 468 • Newton, NC 28658-0468

Instruments
and Disposables
for Medicine
and Science

Indications for Use Statement

510(k) Number: **K032150**

Device Name: Safety-Multifly®

Indications For Use:

The Safety-Multifly® is a winged blood collection needle set with flexible tubing, an adapter for the S-Monovette® system, and an integral needle protector. The Safety-Multifly® is intended for blood collection via venipuncture. The Safety-Multifly® needle protector, when activated after removal from the patient, covers the used needle until it can be properly disposed; therefore, aiding in the prevention of needle stick injuries.

The Safety-Multifly® is individually packaged, sterile, single-use and non-pyrogenic.

The Safety-Multifly® is intended to be used in all hospitals and/or practices where S-Monovettes are used.

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Prescription Use ✓
(Per 21 CFR 801.109)

OR

Over-The-Counter Use _____

Rafaela Cuente

(Division Sign-Off)
Division of Anesthesiology, General Hospital,
Infection Control, Dental Devices

510(k) Number: K032150

Sarstedt, Inc.
P.O. Box 468
Newton, NC 28658-0468

Phone (828) 465-4000 Corporate
Telefax (828) 465-0718 Corporate
Phone (800) 257-5101 Sales
Telefax (828) 465-4003 Sales